



Digital Test Gauge

Operating Instructions




Congratulations on your purchase of the Ashcroft® digital test gauge with total error band full-scale accuracy and the largest display readout in the industry of .66" high. Other industry-leading features include twelve selectable engineering units, seven languages, and password-protected disable and calibration functions. With the range printed on the keypad, the Ashcroft digital gauge meets the ASME B40.7 digital gauge specification.

See a complete listing of product features and specifications on pages 14 & 15.


TABLE OF CONTENTS

	<i>Page</i>
Quick Reference Guide	4-5
Keypad Functions	6-7
• ON/OFF KEY	
• BACKLITE KEY	
• MIN/MAX KEY	
• ZERO/CLEAR KEY	
• ENTER KEY	
• CONFIG KEY	
Configurable Functions (CONFIG Mode)	7-12
• Units (Engineering)	7
• Update Rate	7-8
• Auto Off	8
• Backlite	8-9
• Languages	9
• Damping	9-10
• Contrast	10
• Calibrate (Gauge Calibration)	10-11
• Disable	12
Available Ranges	13
Specifications	14-15
Installation & Battery Replacement	16
• Mounting	
• Battery Replacement & Installation	

QUICK REFERENCE



RENCE



AUTO OFF
(Turns unit off after option selected)

Never*
2 minutes
5 minutes
15 minutes
30 minutes

CALIBRATE
Zero and span adjustments, password protected

CONTRAST
(Customizes display readout)
7 available

Selection 4 is Default*

DISABLE
Allows for "lockout" of CONFIG options

***Indicates Default**

KEYPAD FUNCTIONS



Turns the gauge on and off. When pressing the ON/OFF key while in the off position, gauge start-up display first indicates the software version followed by the model number and gauge pressure range. The gauge will then display indicated pressure and be ready for use.



Manually turns backlight on and off. (See **CONFIG** mode for options).



Allows review of minimum and maximum pressure values since unit start-up or last push of the clear button. Press key to:

- 1) Indicate maximum pressure.
- 2) Indicate minimum pressure.
- 3) Exit MAX/MIN mode and return to unit of pressure measurement mode. To clear minimum and maximum values press ZERO/CLR button. Must be in MAX/MIN mode.

Note: MIN/MAX data is lost when unit is turned off.



Press this key prior to gauge usage to rezero any initial offset less than $\pm 5\%$ of the rated gauge range. If indicated pressure is greater than 5% of range, the rezero feature becomes inoperable. This prevents accidental tare of a pressurized gauge.

To clear minimum and maximum values, press ZERO/CLR button (when min/max values are indicated).



Used in conjunction with CONFIG key, see next page.

KEYPAD FUNCTIONS



CONFIG

This key allows for customization of the gauge.

Pressing the CONFIG key allows cycling through the main menu items; UNITS, UPDATE, AUTO OFF, BACKLITE, LANGUAGE, DAMPING, CONTRAST & CALIBRAT.

UNITS: 12 units of measurement are available, both English and metric, by cycling through the UNITS key; psi, "Hg, "H₂O (with three temperature options, 60°F, 4°C and 20°C), ftSw, Bar, mBar, kPa, mPa, mmHg, cmH₂O, mmH₂O, kg/cm².

Step 1: Press the CONFIG key until the word UNITS appears.

Step 2: Press the ENTER key.

Step 3: Press the CONFIG key once to select ENGLISH or again to select METRIC.

Step 4: Press the ENTER key with selection of ENGLISH or METRIC.

Step 5: Press CONFIG key to select unit of measurement.

Step 6: Press ENTER key to finalize unit of measurement.

UPDATE: this option allows for changing the rate at which pressure is updated on the display screen. The default rate measures pressure at the maximum rate of **10** updates per second or 100 milli-seconds. Optional rates of measurement are measured in updates per second. The options are **10**, 5, 2 or 1 update of pressure measurement per second.

*Indicates default.

KEYPAD FUNCTIONS

Since customer processes vary, update rates should be selected based on the application.

To use the UPDATE option:

Step 1: Press the CONFIG key until the word UPDATE appears.

Step 2: Press the ENTER key.

Step 3: Press the CONFIG key to select the desired update rate.

Step 4: Press ENTER key to finalize **UPDATE** rate.

AUTO OFF: this option sets the amount of time before the gauge will turn itself off after no activity. Offerings are **Never***, 2, 5, 15, or 30 minutes.

To use the AUTO OFF option:

Step 1: Press the CONFIG key until the word AUTO OFF appears.

Step 2: Press the ENTER key.

Step 3: Press the CONFIG key to select the desired AUTO OFF rate.

Step 4 Press the ENTER key to finalize the AUTO OFF rate.

BACKLITE: 5 options are available. They include **ON/OFF***, 10 seconds, 30 seconds 1 or 5 minutes. With the ON option pressed, the gauge backlite will remain lit whenever the gauge is in the ON mode or until the backlite button is pushed again. Options allow the backlite to automatically turn-off after a selected period of time. **Note:** leaving backlite button on will decrease battery life.

*Indicates default.

KEYPAD FUNCTIONS

To use the **BACKLITE** option:

Step 1: Press the CONFIG key until the word BACKLITE appears.

Step 2: Press the ENTER key.

Step 3: Press the CONFIG key to select one of the available time options.

Step 4: Press the ENTER key to finalize your choice of BACKLITE options.

LANGUAGE: available in seven different languages, this option allows the user to change the default language in the CONFIG mode. The languages include **English***, French, Spanish, German, Italian, Portuguese and Dutch.

Step 1: Press the CONFIG key until the word LANGUAGE appears.

Step 2: Press the Enter key.

Step 3: Press the CONFIG key to select one of the available LANGUAGE options.

Step 4: Press the ENTER key to finalize your LANGUAGE option.

DAMPING: with six different options, this mode allows for taking process pressure readings and averaging them. This option is particularly useful when there is pulsation in the process. The options are **NONE***, AVG 2, 4, 6 or 8.

Step 1: Press the CONFIG key until the word DAMPING appears.

Step 2: Press the ENTER key.

Step 3: Press the CONFIG key to select one of the available DAMPING options.

*Indicates default.

KEYPAD FUNCTIONS

Step 4: Press the ENTER key to finalize your DAMPING option.

CONTRAST: this mode allows for BACKLITE contrast level. Seven options are available, 1, 2, 3, 4*, 5, 6 and 7.

Step 1: Press the CONFIG key until the word CONTRAST appears.

Step 2: Press the ENTER key.

Step 3: Press the CONFIG key to select one of the available CONTRAST options.

Step 4: Press the ENTER key to finalize your CONTRAST selection.

Note: setting high contrast levels will decrease battery life.

CALIBRAT.:

Gauge Calibration: Both zero and span adjustments are available. This gauge has been configured with a default password of 00000. This factory password does not allow access to calibration. To access the calibration mode, it is necessary to configure a *user password*. Once the user password is configured, it will become the default password that allows access to gauge calibration.

To access the factory default password:

Step 1: Press the CONFIG key until the word CALIBRAT appears.

Step 2: Press the ENTER key.

Step 3: The letters/astericks... PW***** appear.

*Indicates default.

KEYPAD FUNCTIONS

Step 4: Press the CONFIG key. An 0 appears in the first position.

Step 5: Press the ENTER key once.

Step 6: Press the CONFIG key until 0 appears. 0 will appear in the second position.

Step 7: Press ENTER.

Step 8: Use this format until all the asterisks are replaced with 0.

There now should be a total of five 0's on the keyboard display. The zero in the fifth position should be blinking.

Step 9: Press the ENTER key. You are now prompted to SET PW (or set password).

Step 10: Press the ENTER key.

Step 11: Decide on a five number user password, then follow the procedure above inserting a number in the flashing display until all five numbers are inserted.

Step 12: A SAVE prompt will then appear. If the selected user password is acceptable, press ENTER. If the selected user password is not acceptable press ZERO CLEAR to refigure the user password.

After the password is configured, the default factory password will be replaced with the user password. Once configured, the factory password is no longer accessible.

If an incorrect password is entered, the system will display WRONG. Press the CONFIG key to reenter the correct password.

Step 13: Press ENTER again to begin calibration.

Note: Calibration feature allows recalibration of zero and span.

*Indicates default.

Step 14: Once the correct password is entered, the user will be prompted to set the pressure to zero followed by setting maximum gauge pressure (must be within 5% of the range).

If the user password is lost or stolen, contact Dresser Instrument customer service at (203) 378-8281 for a new factory password that will allow the user to establish a new user password.

DISABLE: allows “lockout” of *individual* CONFIG options. The default is ENABLE for all options in the CONFIG mode.

Step 1: Press the CONFIG key until the word DISABLE appears.

Step 2: Press the ENTER key.

Step 3: Insert the *user* password following the procedure as described in steps 3 through 13 above.

Note: This is the same user password as in the CALIBRAT mode.

Step 4: Press the ENTER key.

Step 5: the first option in the CONFIG menu will now be displayed (UNITS).

Step 6: To DISABLE the UNITS option press the ENTER key until the word DISABLE appears.

Step 7: Press the CONFIG key. The UNITS option is now DISABLED.

Step 8: Proceed through the balance of the CONFIG menu options by pressing the CONFIG key. Follow steps 6-8.

DIGITAL TEST GAUGE RANGES:

psi	Compound (psi)	kPa	Bar/ KSC	Compound (bar)
vac.	15# & vac.	25	1	-1 to 3
10	30# & vac.	40	1.6	-1 to 30
15		60	2.5	
30		100	4	
60		160	6	
100		250	10	
160		400	16	
200		600	25	
300		1000	40	
500			60	
600			100	
800			160	
1000			250	
1500			400	
2000			500	
2500				
3000				
5000				
7500				

mmH ₂ O	mPa	mBar/ cmH ₂ O	Absolute (psia)
3000	1	250	15
5000	1.6	300	25
10,000	2.5	400	50
	6	500	
	100	600	
	160	1000	
	250	1600	
	400	2000	
	500	2500	
		4000	
		5000	
		6000	
		10,000	

SPECIFICATIONS

Type	2089 (0.05% accuracy), 2086 (0.10% accuracy), 2084 (0.25% accuracy)
Accuracy	0.05%, 0.10%, 0.25% all Full Scale Terminal Point Total Error Band (TEB) Accuracy Including Hysteresis, Linearity, Repeatability & Temperature (–18/65°C) (0/150°F)
Dial Size	3"
Case Material	300 Series SS
Case Finish	Electropolished
Case Enclosure Rating	Weatherproof, IP65
Socket Material	316 SS
Socket Size	¼ NPT Male (others on application)
Connection Location	Lower, 3:00, 9:00
Ranges	Vac thru 7500 psi (see engineering units below for other units of measurement)
Operating Temperature	0/150°F
Storage Temperature	–40/180°F
DISPLAY:	
Type	LCD
Display Digits	4½
Character Height	.66"
Backlite	Off By Default
Bar Graph	Yes
Battery Life	1000 Hrs.
Agency Approvals	CE, FM (CSA, Cenelec pending)
KEYPAD FUNCTIONS:	
On/Off	Manually Turns Unit On & Off (auto off options in config menu)
Backlit	Manually Turns Backlit On & Off (auto off options in config menu)
Min/Max	Stores Min & Max Values
Zero/Clear	Zeros Display or Clears Min/Max Values When Displayed

SPECIFICATIONS

Enter	Selects Items In CONFIG Menu
Config Mode	Allows Scrolling Through CONFIG Menus
Engineering Units	Psi, "Hg, "H ₂ O (with three temperature options, 60°F, 4°C and 20°C), ft.SW, bar, mbar, kPa, mPa, mmHg, cmH ₂ O, mmH ₂ O, kg/cm ²
Update Rate	4 Options: 10x/sec, 5x/sec, 2x/sec, 1x/sec
Auto Off	6 Options: Never, 2 Min., 5 Min., 15 Min., 30 Min.,
Dampening	6 Options: None, Average 2, 4, 6, 8 x update rate
Languages	7 Languages: English, Spanish, French, Italian, German, Portugese, Dutch
Backlite	5 Options: On/Off, 10 Sec., 30 Sec., 5 Min., 15 Min.
Field Recalibration	Zero & Span (password protected)
Contrast	7 Available Options
Disable Config Options	Allows disabling of Config Options (password protected)
Calibration Chart	10 Point Individual NIST Traceable Calibration Chart, Standard
Standard Accessories	300 Series SS Protective Cover Nylon Protective Carrying Pouch

WARNING AND ERROR MESSAGES

Display	Description
Flashing 0% or 100%	Gauge over-/underpressured beyond 105% of range
LOW BAT	Low battery, replace
ERROR	Internal error, call customer service (203) 378-8281
RES ERROR	Pressure unit conversion exceeds display resolution or gauge pressured beyond resolution

Gauge Installation:

The Ashcroft digital test gauge comes standard with a 1/4 NPT connection. Good piping practices recommend using teflon tape or a pipe sealant on the gauge threads. Utilize a 7/16" wrench on the wrench flat of the gauge to tighten the gauge to the process.

NEVER TIGHTEN GAUGE THREADS BY HOLDING THE BODY OF THE GAUGE. DOING SO MAY DAMAGE THE GAUGE AND MAKE THE GAUGE INOPERABLE.

Battery Installation & Replacement:

The gauge comes standard with a quantity of three AAA alkaline batteries (installed).

Batteries have a life of approximately 1000 hours. Battery life is dependent on gauge usage, backlite settings and power off settings. When the display flashes LOW BAT, batteries should be replaced.

To replace the batteries:

- 1) Remove the single screw on the back of the gauge case.
- 2) Hold the keypad in the palm of hand.
- 3) Carefully remove the three batteries from the holder and replace the batteries. Use only AAA alkaline non-rechargeable batteries. Do not mix ages or brands of batteries. Do not replace batteries in hazardous areas.



Pipe to which gauge is attached must be properly grounded.



Dresser Instrument
Dresser, Inc.
250 Main Street
Stratford, CT 06614-5145
Tel: (203) 378-8281
Fax: (203) 385-0602
e-mail: info@dresser.com